



**EC Declaration of Conformity**

Kinetik Medical Devices Ltd, herewith declare that the products

**Product name: Infrared Ear Thermometer**  
**Type: ET-100D**  
**UMDNS-Code: [17-887]**

meets the provisions of Directive 93/42/EEC which apply to them.

The medical devices has been assigned to class IIa according to Rule of section III of Annex IX of the Directive 93/42/EEC. It bears the mark

**C € 0197**

The product concerned has been designed and manufactured under a quality management system according to Annex II (excluding section 4) of Directive 93/42/EEC.

Compliance of the designated product with the Directive 93/42/EEC has been assessed and certified by the Notified Body

**TÜV Rheinland LGA Products GmbH**  
**Tillystraße 2, 90431, Nürnberg, Germany**

Issue date: 22.02.2019  
Expiry date: 16/02/2024

following the procedure relating to the EC Declaration of Conformity set out in Annex II of Directive 93/42/EEC.

This Declaration of Conformity covers all medical devices as specified in the product list belonging to this declaraton and is only valid in connection with the batch specific Certificate of Compliance for all products concerned bearing the CE mark.

Place / Date: Surrey / November 3, 2020

Signed: *JRGrover*  
Name: James Grover  
Position: Sales Director