

In association with

St John
Ambulance





kinetik
WELLBEING

-Quick Start Guide-

AG-607

Blood Glucose Monitoring System

1




Wash and dry hands

—

Using warm water
may help
the blood flow.

2




Insert test strip into strip port with contact bars facing toward you.

Ensure that the device is off when inserting the test strip.


The device will automatically turn on and be ready for testing once test strip has been inserted.

3



Prepare the lancing device. Ensuring to insert the lancet firmly into lancing cup holder and set the lancet level appropriately.

4



Obtain a blood sample.


Hold lancing device
against the site
to be lanced and then
press the release button
to puncture.

5



Squeeze lancet site until a drop of blood forms. Wipe away the first blood drop and then squeeze until a second small blood drops forms.


6



Hold the device at a 45° angle.

Then gently apply the blood sample to the underside of the test strip at the bottom - where the absorbent hole is located.

7



Remove lancet site from the test strip when the 5 second countdown begins. The test result will appear on meter after countdown.

The device will turn off 30 seconds after removing test strip.



for more information visit :
kinetikwellbeing.com