


Guidance and manufacturer's declaration - electromagnetic immunity

The IET2 Device is intended for use in the electromagnetic environment specified below. The customer or the user of the IET2 Device should assure that it is used in such an environment.			
IMMUNITY test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Radiated RF IEC 61000-4-3	3 V/m 80 MHz to 2.5 Ghz	3 A/m	Portable and mobile RF communications equipment should be used no closer to any part of the IET2 Device, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter. Recommended separation distance : $d = 1.2\sqrt{P}$ 80 MHz to 800 MHz $d = 2.3\sqrt{P}$ 800 MHz to 2,5 GHz Where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in meters (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey, <sup>a</sup> should be less than the compliance level in each frequency range. <sup>b</sup> Interference may occur in the vicinity of equipment marked with the following symbol: 
NOTE 1: At 80 MHz and 800 MHz, the higher frequency range applies.			
NOTE 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			
A. Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the IET2 Device is used exceeds the applicable RF compliance level above, the IET2 Device should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the IET2 Device. B. Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m.			

Recommended separation distances between portable and mobile RF communications equipment and the IET2 Device

The IET2 Device is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the IET2 Device can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the IET2 Device as recommended below, according to the maximum output power of the communications equipment.			
Rated maximum output power of transmitter W	Separation distance according to frequency of transmitter		
	150 kHz to 80 Mhz $d = 1.2\sqrt{P}$	80 MHz to 800 Mhz $d = 1.2\sqrt{P}$	800 MHz to 2,5 Ghz $d = 2.3\sqrt{P}$
0.01	0.12	0.12	0.23
0.1	0.38	0.38	0.73
1	1.2	1.2	2.3
10	3.8	3.8	7.3
100	12	12	23

# Inner Ear Thermometer

## User Instructions

### Please read the instructions carefully before using.

The ear thermometer is a device capable of achieving infra-red temperature measurement when placed in the auditory canal of a subject. It is a safe means of measuring human body temperature through the ear.

#### Specification:

The standard we adopted is EN12470-5:2003 Clinical thermometers -- Part 5: Performance of infrared ear thermometers (with maximum device).  
This device complies with the requirements of IEC 60601-1, IEC 60601-1-11 and IEC 60601-1-2.

**Range of displayed temperature :** 34°C~44°C(93.2°F~111.2°F)

Display L°C(L°F) when the temperature is under 34°C(93.2°F)

Display H°C(H°F) when the temperature is over 44°C(111.2°F)

#### Operating ambient:

Temperature: 16°C~35°C(60.8°F~95.0°F)

Humidity: ≤80%rh

#### Storage and transport condition:

Temperature: -10°C to 55°C(14°F to 131°F)

Humidity: 30%rh to 90%rh

**Display resolution temperature range:** 0.1°C(0.1°F)

**Accuracy:** ±0.2°C (from 35.5°C to 42.0°C)

±0.4°F (from 96.0°F to 107.6°F)

#### Display:

liquid crystal display, 4 digits

temperature value: display the maximum temperature

in measuring process

temperature unit: centigrade or fahrenheit

display of memory: last ten memories

low battery warning: the LCD display  and then a beep sound is heard

**Power consumption:** ≤ 30 mW

**Battery:** two 1.5V alkaline batteries (AAA)

**Battery life:** 4000 takes or 6 months with 3-min. use per day

**Dimension:** 134mm by 37mm by 67mm


**Net weight:** 74g

**Beeper sign:** on/off, measuring finish and low voltage warning, etc.



**Self-testing sequence:** Press the 'on/recall' button to turn on the thermometer and all of symbol (See Fig A) should be displayed on the LCD in one second.

**Calibration:** This thermometer is calibrated at the time of manufacture. If the thermometer is used according to the instructions, periodic re-calibration is not required.

**Manufacturing Batch code:** see the label within the battery compartment.




### Cautions

1. Condition the thermometer at room temperature for 30 minutes before taking a reading.
2. Use a new probe cover with each and every measurement to prevent inaccurate readings.
3. Please keep your ear canal clear, if not, the measuring result shall be inaccurate.
4. The ' ? ' symbol on the display indicates the device is ready for use.
5. Do not scratch the probe membrane as this will affect the accuracy of measurements.
6. If the LCD does not change or respond when you press any of the buttons, remove the batteries and then replace to reset.
7. Please don't use the thermometer if your ear canal has become inflamed. Do not force the thermometer into the ear canal. If pain occurs, stop using immediately.
8. Do not expose this thermometer to electric shock.
9. Do not expose the thermometer to sunlight or to water.
10. Do not use near strong electromagnetic fields, i.e. Keep it away from any radio systems and mobile phones.
11. Do not modify this device without authorization of the manufacturer.
12. This device must always be kept in a clean, dry area.
13. : Degree of protection against electric shock is Type BF applied part.
14. : Battery disposal should be completed in accordance with local regulations.


**Note:**

- In order to avoid the spreading of germs, use a new probe cover for each measurement
- Check to ensure probe cover is fitted on firmly before use (please see the diagrams right).  
If the probe cover is broken, discard and apply a new one immediately.




**How to use (°C displays for example )  
Measuring human body temperature in the Ear Canal**

1. Press the 'on/recall' button to turn on the thermometer. After the beep the thermometer will be ready to read when the LCD display matches Fig B.



2. Straighten the ear canal by pulling the outer ear up and back to give a clear view of the eardrum.


- For children under 1 year, Pull the ear straight back.
- For children ages 1 year to adult, Pull the ear up and back.  
(Caution: Many pediatricians suggest use of ear thermometers for children older than 6 months.)




3. While pulling the outer ear, insert the probe snugly into the ear canal (insert as deep as is comfortable, do not force, the probe into the ear canal). Press and immediately release the 'SCAN' button. Do not remove the thermometer from the ear until you hear the beep sound – this beep means that the measurement is complete. Remove the thermometer from the ear. The display shows the measured temperature.

4. **Repeat measurements:** wait at least 1 minute between measurements. Ensure the Ear symbol is visible in the LCD.

5. **Memory seach:** Press the 'on/recall' button, the last ten memories (NO.9-NO.0) will take turns to display on the LCD as Fig C and Fig D.



6. The LCD will displays 'Er 1' as Fig E and cannot displays temperature when the ambient temperature is out of the range of 16°Cto 35°C.



**Note: Before measuring, the thermometer shall be stabilized at the operating ambient condition for a minimum of 30 min.**

7. **Shut off:**The thermometer shall automatically shut off without manipulating in one minute.

**Changing from Celsius to Fahrenheit**

To switch the display between °C and °F, turn the unit on. Press and hold the on/recall button for 3 seconds, the display will show as Fig. F, the unit will prepare for a measurement.



**Explanation of symbols**

battery is empty	Model Reference	or  Read IFU carefully
Product disposal instructions for electronic devices	Memory	Authorised representative in the european community
The battery in this product complies with the requirements stated in European Directives 2006/66/EEC	Type BF equipment	Symbol for "MANUFACTURER"
	Symbol for "COMPLIES WITH MDD 93/42/EEC REQUIREMENTS"	Classification according to the degree of protection against ingress of water as detailed in IEC 529
	LC or LF subject temperature is under 34°C	Lot Number
	HC or HF subject temperature is above 44°C	

**Product disposal**


Please ensure environmental protection. Batteries do not belong in the domestic waste. Please hand them in at collection point or the municipal recycle material centre as special waste.

This symbol on products and/or accompanying documents means that consumed electronic products must not be mixed with conventional domestic waste. Take these products to the corresponding collection points for correct treatment and recycling, where they will be accepted free of charge. For more information on the closest collection point, Please enquire with your local authorities.



**Cleaning and Storage**

1. Store thermometer and the **Disposable probe covers** in a dry location free from dust and contamination and away from direct sunlight. The ambient temperature at the storage location should remain fairly constant and within the range of -10°C to 55°C.
2. Use an alcohol swab or cotton swab moistened with alcohol (70% Isopropyl) to clean the thermometer casing and the measuring probe before each use. Ensure that no liquid enters the interior of the thermometer.
3. Never use abrasive cleaning agents, thinners or gasoline for cleaning and never immerse the instrument in water or other cleaning liquids. Take care not to scratch the surface of the probe membrane or display.



**Replacing the batteries**

1. When voltage of the batteries are low, the LCD will display ' ' symbol, please replace two new batteries in unit. The thermometer cannot work accurately under the condition of low voltage.
2. The thermometer is supplied with two 1.5V ALKALINE BATTERIES (AAA). Insert new batteries when the low battery symbol appears on the LCD.
3. Remove the battery cover and take out the old batteries.
4. Place two new batteries according to the '+' or '-'
5. Please take out the batteries to avoid battery leaking if unit not used for over six months.
6. The disposal of the probe cover, battery and device shall comply with the local environment requirements.  
The lithium battery or fuel cell may lead to excessive temperatures, fire or explosion.

**Guarantee**

This product is guaranteed for a period of one year from the date of purchase against mechanical and electrical manufacturing defects. There are no serviceable parts inside this device. Any attempted repair by unauthorised persons invalidates the warranty. In the unlikely event that you experience a problem, please return it to the retailer where you made the purchase, along with your receipt. This does not affect your statutory rights.

**Guidance and manufacturer's declaration - electromagnetic emissions**

The IET2 Device is intended for use in the electromagnetic environment specified below. The customer or the user of the IET2 Device should assure that it is used in such an environment.		
Emissions test	Compliance	Electromagnetic environment - guidance
RF emissions CISPR 11	Group 1	The IET2 uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions CISPR 11	Class B	The IET2 is suitable for use in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
Harmonic emissions IEC 61000-3-2	Not applicable	
Voltage fluctuations/flicker emissions IEC 61000-3-3	Not applicable	

**Guidance and manufacturer's declaration - electromagnetic immunity**

The IET2 Device is intended for use in the electromagnetic environment specified below. The customer or the user of the IET2 Device should assure that it is used in such an environment.			
IMMUNITY test	IEC 60601 test level	Compliance	Electromagnetic environment - guidance
Electrostatic discharge (ESD) IEC 61000-4-2	6 kV contact 8 kV air	6 kV contact 8 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.
Power frequency (50/60 Hz) magnetic field IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.