
*Single Sided Printing


Front of the QSG

50mm

75mm




Unpack device, connect type-C cable and charge fully before use.




Attach cord with two electrode pads.




Remove protective film and attach gel pads to areas of treatment.




Hold down the power button for 2 seconds to turn it on.



Press the power button to choose your desired treatment mode.



Use the + or - button to adjust the intensity and start the treatment.



Press and hold the power button for 2 seconds to switch off.

Scan for support

WW_WW-200_UK_QSG_20250501_v3

54mm

79mm

kinetik WELLBEING

PRECISION PERIOD PAIN RELIEVER

WITH TENS TECHNOLOGY

SKU WW-200

REF AD-2300S

Cover

54mm

79mm

kinetik WELLBEING

For more information visit:
kinetikwellbeing.com

Back Cover